



JOINT EUROPEAN RESEARCH INFRASTRUCTURE NETWORK FOR COASTAL OBSERVATORIES

FIRST CALIBRATION EXPERIMENT AND PERSPECTIVES

Ifremer Brest – Metrology laboratory

F. Salvetat | Ifremer | florence.salvetat@ifremer.fr



Sharing experiences,
practices
through networking

Data Assurance Quality



Metrology

Jerico WP4



IFREMER'S METROLOGY LABORATORY



The metrology laboratory:

Dedicated to physical and physico-chemical oceanographic parameters:

Temperature (Cofrac), pressure (Cofrac), salinity (conductivity), velocity, ocean current, dissolved oxygen, pH, turbidity, fluorescence.




Salinometer

www.jerico-fp7.eu

Fresh water or seawater bath (800l)



Formazin solutions



Fluorescein solutions



Towing canal



THE EXPERIMENT

Calibration experiment:

- Conductivity
- Temperature
- Oxygen

Fresh water or
seawater bath
(800l)



Fresh water or
seawater oxygen
bath (100l)





THE EXPERIMENT

Temperature /conductivity calibration: (Cofrac in temperature)

- Calibration points

Salinity	Temperature (°C)
35	5
	15
17	15
10	20



THE EXPERIMENT

Temperature /conductivity calibration: (Cofrac in temperature)

- Protocol

Comparison to Standard Platinum Resistance Thermometer and autosal salinometer (both calibrated).

SPRT



Resistance
bridge



Salinometer



THE EXPERIMENT



Temperature /conductivity calibration: (Cofrac in temperature)

- Temperature calibration uncertainty (for Seabird)

$$U = \pm 10 \text{ m}^\circ\text{C}$$



$$U = \pm 40 \text{ m}^\circ\text{C}$$



100L bath
(heterogeneity
less than
10m°C)



800L bath
(heterogeneity=
40m°C)



THE EXPERIMENT

Dissolved oxygen calibration:

- Calibration points

Dissolved Oxygen	Temperature (°C)	Salinity
100% (air equilibrium)	20	0
	10	0
50%	20 or 10	0



THE EXPERIMENT

Dissolved oxygen calibration:

- Protocol

Comparison to Winkler analysis.



Winkler titrator



Winkler samples

First calibration experiment and perspectives JERICO - 9



THE EXPERIMENT

Dissolved oxygen calibration:

- Dissolved oxygen calibration uncertainty

Expectation: $U = \pm 4 \mu\text{mol/L}$

Depend on sensors uncertainty components.



THE EXPERIMENT



The experiment:





The experiment:

Participating laboratories:


- HCMR (Greece): Tanya Tsagaraki, Manolis Ntoumas and George Petihakis
Sbe37-SIP CT sensor and Aandera 3830 DO optode
- CNR-ISMAR (Italy): Stefania Sparnocchia and Elio Paschini
Sbe19 plus CTD sensor and Sbe43 DO sensor
- AZTI Tecnalia (Spain): Carlos Hernandez
Sbe37-SMP CTD sensor
- NIVA (Norway): Emanuele Reggiani,
Aandera 3830 DO optode



THE EXPERIMENT

The experiment:

Conductivity and temperature (in progress)



THE EXPERIMENT



The results:

- Calibration certificates or reports for each institute
 - Inter Laboratory Comparison anonymous report.
 - Calculation of the normalized En scores (indication of laboratories agreement)
- or
- Comparison of the sensors errors

Perspectives

- Testing other parameters (turbidity, fluorescence, ...)
- Testing different sensor technologies (conductivity sensor: inductive or electrodes sensors)
- Comparing different calibration protocols performed in different institutes.
- Performing in field experiments (ACT ?)



THANKS FOR YOR ATTENTION